



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: May 19, 2010 INITIALS: DEC OWNER: WASHINGTON STATE PATROL

DISTRICT: 6 ADDRESS: SR97 MP315

LOCATION Tonasket 97

COUNTY: Okanogan

MAKE: Fairbanks

TYPE Trk Scale

ACCURACY CLASS: IIII

MODEL: SP-CA

ID. NO: G408859

CAPACITY: 50000 lbs. SECTIONS: 2

MINIMUM GRADUATIONS: 50 lbs.

REG. ELEMENT: Dial

PLATFORM 8x10

| LOAD POSITION | TEST WEIGHT | SCALE INDICATION | ERROR | LOAD POSITION | TEST WEIGHT | SCALE INDICATION | ERROR |
|---------------|-------------|------------------|-------|---------------|-------------|------------------|-------|
| CENTER | 4000 | 4000 | 0 | CENTER | 20000 | 20000 | 0 |
| CENTER | 8000 | 8000 | 0 | SEC 1 | 20000 | 20000 | 0 |
| CENTER | 12000 | 12000 | 0 | SEC 2 | 20000 | 20000 | 0 |
| CENTER | 16000 | 16000 | 0 | | | | |

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-1 and 200446-0-L2902-3

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by:

David Cromer

Date:

July 1, 2010

Print name: David Cromer

State of Washington

County of Shureton

Signed and sworn to (or affirmed before
me on July 1, 2010 by

Brenda A Biscay
(Signature of notary public)

Brenda A Biscay

Printed Name of Notary

My appointment expires 3.12.12

